

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KH2002 Environment (Paddle type) Flowmeter, Electromagnetic Flowmeter, Vortex Flowmeter, Ultrasonic Flowmeter, Paddle Wheel Flowmeter, Positive Displacement Flowmeter, Propeller Water Meter, Turbine Flowmeter, Electronic Flowmeter, Mechanical water meter	(1) 0.2 grade Electromagnetic Flowmeter/YOKOGAWA /AXF200G (2) 0.2 grade Electromagnetic Flowmeter/YOKOGAWA /AXF100G (3) 0.2 grade Electromagnetic Flowmeter/YOKOGAWA /AXF050G (4) 0.2 grade Electromagnetic Flowmeter/YOKOGAWA /AXF025G	In-house method: Large pipe flowmeter calibration procedure (Document No.: LAB-C101) Small pipe flowmeter calibration procedure (Document No.: LAB-C102)	2.5	m ³	30	m ³	Using Standard Meter (1) @ (41 to 420) m ³ /h	0.58	%
			1.5	m ³	10	m ³	Using Standard Meter (2) @ (12.5 to 125) m ³ /h	0.52	%
			0.3	m ³	2	m ³	Using Standard Meter (3) @ (4 to 31.5) m ³ /h	0.48	%
			0.1	m ³	0.6	m ³	Using Standard Meter (4) @ (0.63 to 7.88) m ³ /h	0.54	%
Approval Signatory: HSU, Yao-Chia; HUANG, Chia-Hung; LAI, Yu-Hsuan; SU, David									

Note : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.
(Null Below)

